EZEEKLEEN 2.5 HD

4-STAGE ULTRA CLEAN WATER SYSTEM

Model: 8000

PROTOCOL FOR USAGE

When drawing water from the Ezee Kleen 2.5 HD ALWAYS ensure the following protocol is used:

- Tip: Photocopy and laminate or contact Oasis Dental Customer Support line for a free laminated copy of this page and post at the Ezee Kleen 2.5 HD dispensing faucet.
- Water is suitable for use in autoclaves, self contained dental unit water systems, ultrasonic baths, instrument cleaning and film developers
- Collect and store Ezee Kleen 2.5 HD autoclave safe water in clean containers marked "Autoclave Water Only" "No Tap Water" "Keep Lid On" "Store Only 48 Hours"
- Label the Dispensing Faucet "Autoclave Water - Always Check Monitor while Drawing Water"
- Disregard monitor when not dispensing water from the unit (ie. when unit sits idle)
- Check Ezee Kleen 2.5 HD Monitor each time water is drawn
- Regularly inspect the containers for cleanliness
- When storing filled containers ensure caps remain on
- Avoid storing or drawing water in an unclean environment
- Avoid storing the water more than 48 hours
- Ensure the working area around the dispensing faucet and especially the dispensing faucet itself remains clean
- Avoid running equipment that has the potential to create dust or particulate matter in the atmosphere when drawing water
- Use "Patient Scheduler" to schedule maintenance for Ezee Kleen 2.5 HD annually
- See page 11 of user' manual for "Changing your Cartridges – Instructions"
- Failure to comply with manufacturer's "Protocol for Usage" and below maintenance recommendations could result in water not suitable for intended use

Cartridge	8002-FR DI Cartridge with drainline flow restrictor	8003 Pre-Filter	8004 RO Membrane	8005 UV Lamp
Maintenance	Change when monitor reads > 5.00 with water flowing (red light will begin to flash as a secondary indication)	Change Annually	Change Bi-annually (Every Two Years)	Change Annually
Last maintained				
Changed by				
Cartridge Supplier:				
Staff member responsible for maintenance of the Ezee Kleen 2.5 HD:				

Questions? Contact customer support Mon – Fri, PST, 8:00am – 5:00pm.

Oasis DENTAL GROUP INC.

16147 - 8A Avenue, Surrey, BC, V4A 8W8 ● Ph: (800) 338-6693 ● (604) 538-1285 ● Fax: (604) 538 - 1295

© 2009 - 2012 Oasis Dental Group Inc.™